



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Datacolor Technology Suzhou
288 Shengpu Road
Suzhou Jiangsu, China 215021

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AC-1759

Certificate Number



ANAB Approval

Certificate Valid Through: 02/20/2021
Version No. 008 Issued: 12/27/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Datacolor Technology (Suzhou) Co., Ltd

288 Shengpu Road, Suzhou Jiangsu, China 215021

Ferris Ding - Technical Lab Manager

Phone: +86 512 6936 9821

Tommy Zhao - Quality Manager

Phone: +86 512 6258 0338

tzhao@datacolor.com

www.datacolor.com

CALIBRATION

Valid to: February 20, 2021

Certificate Number: AC-1759

Photometry and Radiometry

Table with 4 columns: Parameter/Equipment, Range, Expanded Uncertainty of Measurement (+/-), Reference Standard, Method, and/or Equipment. Rows include Spectral Reflectance Factor for various geometries and Bidirectional Spectral Reflectance Factor.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. The uncertainty term R represents Spectral Reflectance.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1759.

Signature of R. Ding
Vice President